



# Louisiana

## Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

**Policy #** 00673

**Original Effective Date:** 05/15/2019

**Current Effective Date:** 06/14/2021

*Applies to all products administered or underwritten by Blue Cross and Blue Shield of Louisiana and its subsidiary, HMO Louisiana, Inc. (collectively referred to as the “Company”), unless otherwise provided in the applicable contract. Medical technology is constantly evolving, and we reserve the right to review and update Medical Policy periodically.*

### Services Are Considered Investigational

*Coverage is not available for investigational medical treatments or procedures, drugs, devices or biological products.*

Based on review of available data, the Company considers the use of a powered exoskeleton for ambulation in patients with lower-limb disabilities to be **investigational**.\*

### Background/Overview

An exoskeleton is an external structure with joints and links that might be regarded as wearable robots designed around the shape and function of the human body. A powered exoskeleton, as described in this evidence review, consists of an exoskeleton-like framework worn by a person that includes a power source supplying energy for limb movement.

One type of powered lower-limb exoskeleton (eg, ReWalk™, Indego®)† provides user-initiated mobility based on postural information. Standing, walking, sitting, and stair up/down modes are determined by a mode selector on a wristband. ReWalk includes an array of sensors and proprietary algorithms that analyze body movements (eg, tilt of the torso) and manipulate the motorized leg braces. The tilt sensor is used to signal the onboard computer when to take the next step. Patients using the powered exoskeleton must be able to use their hands and shoulders with forearm crutches or a walker to maintain balance. Instructions for ambulating with ReWalk are to place the crutches ahead of the body, and then bend the elbows slightly, shifting weight toward the front leg, leaning toward the front leg side. The rear leg will lift slightly off of the ground and then begin to move forward. Using the crutches to straighten up will enable the rear leg to continue moving forward. The process is repeated with the other leg.

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.



# Louisiana

## Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

To move from a seated to standing position or vice versa, the desired movement is selected by the mode selector on the wrist. There is a 5-second delay to allow the individual to shift weight (forward for sit-to-stand and slightly backward for stand-to-sit) and to place their crutches in the correct position. If the user is not in an appropriate position, a safety mechanism will be triggered. Walking can only be enabled while standing, and the weight shift must be sufficient to move the tilt sensor and offload the back leg to allow it to swing forward. Continuous ambulation is accomplished by uninterrupted shifting onto the contralateral leg. The device can be switched to standing either via the mode selector or by not shifting weight laterally for 2 seconds, which triggers the safety mechanism to stop walking. Some patients have become proficient with ReWalk by the third week of training.

## **FDA or Other Governmental Regulatory Approval**

### **U.S. Food and Drug Administration (FDA)**

In 2014, ReWalk (ReWalk Robotics, previously Argo Medical Technologies) was granted a de novo 510(k) classification (K131798) by the FDA (Class II; FDA product code: PHL). The new classification applies to this device and substantially equivalent devices of this generic type. ReWalk is the first external, powered, motorized orthosis (powered exoskeleton) used for medical purposes that is placed over a person's paralyzed or weakened limbs for the purpose of providing ambulation. De novo classification allows novel products with moderate- or low-risk profiles and without predicates that would ordinarily require premarket approval as a Class III device to be down-classified in an expedited manner and brought to market with a special control as a Class II device.

The ReWalk is intended to enable individuals with spinal cord injury at levels T7 to L5 to perform ambulatory functions with supervision of a specially trained companion in accordance with the user assessment and training certification program. The device is also intended to enable individuals with spinal cord injury at levels T4 to T6 to perform ambulatory functions in rehabilitation institutions in accordance with the user assessment and training certification program. The ReWalk is not intended for sports or stair climbing.

Candidates for the device should have the following characteristics:

- Hands and shoulders can support crutches or a walker
- Healthy bone density

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.



# Louisiana

## Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

- Skeleton does not suffer from any fractures
- Able to stand using a device such as a standing frame
- In general good health
- Height is between 160 cm and 190 cm (5'3" to 6'2")
- Weight does not exceed 100 kg (220 lb).

In 2019, the ReWalk ReStore<sup>TM†</sup>, a lightweight, wearable, exo-suit, was approved for rehabilitation of individuals with lower limb disabilities due to stroke.

In 2016, Indego (Parker Hannifin) was cleared for marketing by the FDA through the 510(k) process (K152416). The FDA determined that this device was substantially equivalent to existing devices, citing ReWalk as a predicate device. Indego is “intended to enable individuals with spinal cord injury at levels T7 to L5 to perform ambulatory functions with supervision of a specially trained companion.” Indego has also received marketing clearance for use in rehabilitation institutions.

In 2016, Ekso<sup>TM†</sup> and Ekso GT<sup>TM†</sup> (Ekso Bionics<sup>®†</sup> Inc) were cleared for marketing by the FDA through the 510(k) process (K143690). The ReWalk was the predicate device. Ekso is intended to perform ambulatory functions in rehabilitation institutions under the supervision of a trained physical therapist for the following populations with upper extremity motor function of at least 4/5 in both arms: individuals with hemiplegia due to stroke; individuals with spinal cord injuries at levels T4 to L5; individuals with spinal cord injuries at levels of C7 to T3.

In 2017, HAL for Medical Use (Lower Limb Type) (CYBERDYNE Inc.) was cleared for marketing by the FDA through the 510(k) process (K171909). The ReWalk was the predicate device. The HAL is intended to be used inside medical facilities while under trained medical supervision for individuals with spinal cord injury at levels C4 to L5 (ASIA C, ASIA D) and T11 to L5 (ASIA A with Zones of Partial Preservation, ASIA B)

In 2020, Keeogo<sup>TM†</sup> (B-Temia) exoskeleton was cleared for marketing by the FDA through the 510(k) process (K201539). The Honda Walking Assist Device was the predicate device. Keeogo is intended for use in stroke patients in rehabilitation settings.

FDA product code: PHL.

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.



# Louisiana

Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

## **Rationale/Source**

### **Description**

The goal of the powered exoskeleton is to enable people who do not have volitional movement of their lower extremities to be able to fully bear weight while standing, to walk, and to navigate stairs. The devices have the potential to restore mobility and, thus, might improve functional status, quality of life, and health status for patients with spinal cord injury, multiple sclerosis, amyotrophic lateral sclerosis, Guillain-Barré syndrome, and spina bifida.

### **Summary of Evidence**

For individuals who have lower-limb disabilities who receive a powered exoskeleton, the evidence includes 1 randomized cross-over study and several case series. Relevant outcomes are functional outcomes, quality of life, and treatment-related morbidity. At the present, evaluation of exoskeletons is limited to small studies primarily performed in institutional settings with patients who have spinal cord injury. These studies have assessed the user's ability to perform, under close supervision, standard tasks such as the Timed Up & Go test, 6-minute walk test, and 10-meter walk test. One randomized, open-label cross-over study and a case series in patients with multiple sclerosis and spinal cord injury, respectively, assessed use of powered exoskeletons in the outpatient setting. Although these small studies indicate powered exoskeletons may be used safely in the outpatient setting, these devices require significant training, and their efficacy has been minimally evaluated. Further evaluation of users' safety with these devices under regular conditions, including the potential to trip and fall should be assessed. Further study is needed to determine the benefits of these devices outside of the institutional setting. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

## **Supplemental Information**

The purpose of the following information is to provide reference material. Inclusion does not imply endorsement or alignment with the evidence review conclusions.

### **Practice Guidelines and Position Statements**

Guidelines or position statements will be considered for inclusion in 'Supplemental Information' if they were issued by, or jointly by, a US professional society, an international society with US representation, or National Institute for Health and Care Excellence (NICE). Priority will be given

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.



Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

to guidelines that are informed by a systematic review, include strength of evidence ratings, and include a description of management of conflict of interest.

**American Physical Therapy Association**

The American Physical Therapy Association published guidelines in 2020 providing recommendations to guide improvement of locomotor function after brain injury, stroke, or incomplete spinal cord injury in ambulatory patients. The guidelines recommend against the use of powered exoskeletons for use on a treadmill or elliptical to improve walking speed or distance following acute-onset central nervous system injury in patients more than 6 months post-injury due to minimal benefit and increased costs and time.

**U.S. Preventive Services Task Force Recommendations**

Not applicable.

**Medicare National Coverage**

There is no national coverage determination. In the absence of a national coverage determination, coverage decisions are left to the discretion of local Medicare carriers.

**Ongoing and Unpublished Clinical Trials**

Some currently ongoing trials that might influence this review are listed in Table 1.

**Table 1. Summary of Key Trials**

NCT No.	Trial Name	Planned Enrollment	Completion Date
<i>Ongoing</i>			
NCT01701388	Investigational Study of the Ekso Powered Exoskeleton for Ambulation in Individuals With Spinal Cord Injury (or Similar Neurological Weakness)	40	Sep 2020 (ongoing)
NCT02658656	Powered Exoskeletons in Persons with SCI (PEPSCI)	160	Sep 2021 (ongoing)

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.



# Louisiana

## Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

NCT04221373	Exoskeletal-Assisted Walking in SCI Acute Inpatient Rehabilitation	40	Jul 2022 (recruiting)
Unpublished			
NCT03082898	Mobility and Therapeutic Benefits Resulting From Exoskeleton Use in a Clinical Setting (SC140121 Study 1 and 2)	41	Jun 2020 (completed)

NCT: national clinical trial; SCI: spinal cord injury.

## References

1. Blue Cross and Blue Shield Association, Medical Policy Reference Manual, “Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities”, 1.03.04, April 2021.
2. Zeilig G, Weingarden H, Zwecker M, et al. Safety and tolerance of the ReWalk exoskeleton suit for ambulation by people with complete spinal cord injury: a pilot study. *J Spinal Cord Med.* Mar 2012; 35(2): 96-101. PMID 22333043
3. Asselin P, Knezevic S, Kornfeld S, et al. Heart rate and oxygen demand of powered exoskeleton-assisted walking in persons with paraplegia. *J Rehabil Res Dev.* 2015; 52(2): 147-58. PMID 26230182
4. Lajeunesse V, Vincent C, Routhier F, et al. Exoskeletons' design and usefulness evidence according to a systematic review of lower limb exoskeletons used for functional mobility by people with spinal cord injury. *Disabil Rehabil Assist Technol.* Oct 2016; 11(7): 535-47. PMID 26340538
5. McGibbon CA, Sexton A, Jayaraman A, et al. Evaluation of the Keeogo exoskeleton for assisting ambulatory activities in people with multiple sclerosis: an open-label, randomized, cross-over trial. *J Neuroeng Rehabil.* Dec 12 2018; 15(1): 117. PMID 30541585
6. van Dijsseldonk RB, van Nes IJW, Geurts ACH, et al. Exoskeleton home and community use in people with complete spinal cord injury. *Sci Rep.* Sep 24 2020; 10(1): 15600. PMID 32973244
7. Tefertiller C, Hays K, Jones J, et al. Initial Outcomes from a Multicenter Study Utilizing the Indego Powered Exoskeleton in Spinal Cord Injury. *Top Spinal Cord Inj Rehabil.* 2018; 24(1): 78-85. PMID 29434463
8. Hartigan C, Kandilakis C, Dalley S, et al. Mobility Outcomes Following Five Training Sessions with a Powered Exoskeleton. *Top Spinal Cord Inj Rehabil.* 2015; 21(2): 93-9. PMID 26364278

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.



# Louisiana

## Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

9. Bach Baunsgaard C, Vig Nissen U, Katrin Brust A, et al. Gait training after spinal cord injury: safety, feasibility and gait function following 8 weeks of training with the exoskeletons from Ekso Bionics. *Spinal Cord*. Feb 2018; 56(2): 106-116. PMID 29105657
10. Esquenazi A, Talaty M, Packel A, et al. The ReWalk powered exoskeleton to restore ambulatory function to individuals with thoracic-level motor-complete spinal cord injury. *Am J Phys Med Rehabil*. Nov 2012; 91(11): 911-21. PMID 23085703
11. Hornby TG, Reisman DS, Ward IG, et al. Clinical Practice Guideline to Improve Locomotor Function Following Chronic Stroke, Incomplete Spinal Cord Injury, and Brain Injury. *J Neurol Phys Ther*. Jan 2020; 44(1): 49-100. PMID 31834165

### **Policy History**

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

- 05/02/2019 Medical Policy Committee review
- 05/15/2019 Medical Policy Implementation Committee approval. New policy.
- 05/07/2020 Medical Policy Committee review
- 05/13/2020 Medical Policy Implementation Committee approval. Coverage eligibility unchanged.
- 09/22/2020 Coding update
- 05/06/2021 Medical Policy Committee review
- 05/12/2021 Medical Policy Implementation Committee approval. Coverage eligibility unchanged.

Next Scheduled Review Date: 05/2022

### **Coding**

*The five character codes included in the Blue Cross Blue Shield of Louisiana Medical Policy Coverage Guidelines are obtained from Current Procedural Terminology (CPT®)‡, copyright 2020 by the American Medical Association (AMA). CPT is developed by the AMA as a listing of descriptive terms and five character identifying codes and modifiers for reporting medical services and procedures performed by physician.*

*The responsibility for the content of Blue Cross Blue Shield of Louisiana Medical Policy Coverage Guidelines is with Blue Cross and Blue Shield of Louisiana and no endorsement by the AMA is*

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.



# Louisiana

## Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

*intended or should be implied. The AMA disclaims responsibility for any consequences or liability attributable or related to any use, nonuse or interpretation of information contained in Blue Cross Blue Shield of Louisiana Medical Policy Coverage Guidelines. Fee schedules, relative value units, conversion factors and/or related components are not assigned by the AMA, are not part of CPT, and the AMA is not recommending their use. The AMA does not directly or indirectly practice medicine or dispense medical services. The AMA assumes no liability for data contained or not contained herein. Any use of CPT outside of Blue Cross Blue Shield of Louisiana Medical Policy Coverage Guidelines should refer to the most current Current Procedural Terminology which contains the complete and most current listing of CPT codes and descriptive terms. Applicable FARS/DFARS apply.*

CPT is a registered trademark of the American Medical Association.

Codes used to identify services associated with this policy may include (but may not be limited to) the following:

Code Type	Code
CPT	No codes
HCPCS	K1007, L9900
ICD-10 Diagnosis	All related diagnoses

\*Investigational – A medical treatment, procedure, drug, device, or biological product is Investigational if the effectiveness has not been clearly tested and it has not been incorporated into standard medical practice. Any determination we make that a medical treatment, procedure, drug, device, or biological product is Investigational will be based on a consideration of the following:

- A. Whether the medical treatment, procedure, drug, device, or biological product can be lawfully marketed without approval of the U.S. Food and Drug Administration (FDA) and whether such approval has been granted at the time the medical treatment, procedure, drug, device, or biological product is sought to be furnished; or
- B. Whether the medical treatment, procedure, drug, device, or biological product requires further studies or clinical trials to determine its maximum tolerated dose, toxicity, safety, effectiveness, or effectiveness as compared with the standard means of treatment or diagnosis, must improve health outcomes, according to the consensus of opinion among experts as shown by reliable evidence, including:

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.





# Louisiana

## Powered Exoskeleton for Ambulation in Patients With Lower-Limb Disabilities

Policy # 00673

Original Effective Date: 05/15/2019

Current Effective Date: 06/14/2021

1. Consultation with the Blue Cross and Blue Shield Association technology assessment program (TEC) or other nonaffiliated technology evaluation center(s);
2. Credible scientific evidence published in peer-reviewed medical literature generally recognized by the relevant medical community; or
3. Reference to federal regulations.

**\*\*Medically Necessary (or “Medical Necessity”)** - Health care services, treatment, procedures, equipment, drugs, devices, items or supplies that a Provider, exercising prudent clinical judgment, would provide to a patient for the purpose of preventing, evaluating, diagnosing or treating an illness, injury, disease or its symptoms, and that are:

- A. In accordance with nationally accepted standards of medical practice;
- B. Clinically appropriate, in terms of type, frequency, extent, level of care, site and duration, and considered effective for the patient's illness, injury or disease; and
- C. Not primarily for the personal comfort or convenience of the patient, physician or other health care provider, and not more costly than an alternative service or sequence of services at least as likely to produce equivalent therapeutic or diagnostic results as to the diagnosis or treatment of that patient's illness, injury or disease.

For these purposes, “nationally accepted standards of medical practice” means standards that are based on credible scientific evidence published in peer-reviewed medical literature generally recognized by the relevant medical community, Physician Specialty Society recommendations and the views of Physicians practicing in relevant clinical areas and any other relevant factors.

‡ Indicated trademarks are the registered trademarks of their respective owners.

**NOTICE:** If the Patient’s health insurance contract contains language that differs from the BCBSLA Medical Policy definition noted above, the definition in the health insurance contract will be relied upon for specific coverage determinations.

**NOTICE:** Medical Policies are scientific based opinions, provided solely for coverage and informational purposes. Medical Policies should not be construed to suggest that the Company recommends, advocates, requires, encourages, or discourages any particular treatment, procedure, or service, or any particular course of treatment, procedure, or service.

©2021 Blue Cross and Blue Shield of Louisiana

Blue Cross and Blue Shield of Louisiana is an independent licensee of the Blue Cross and Blue Shield Association and incorporated as Louisiana Health Service & Indemnity Company.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, or otherwise, without permission from Blue Cross and Blue Shield of Louisiana.